

USER MANUAL



GYNAECOLOGY CHAIR

For models
015-2031, 016-2031

NOTE!

**The operator must read the user manual
carefully BEFORE the product is used,
to avoid unnecessary risks to the patient and operator.**

TABLE OF CONTENTS

Foreword, Warranty, Important information

1. Function Description

1.1 General description

2. Operation

2.1 Connection

2.2 Adjusting the height

2.3 When the patients sits down

2.4 Adjusting the backrest

2.5 Adjusting the seat section/Trendelenburg position

2.6 Replacing the paper roll

2.7 The function of the castors

2.8 Pull-out leg section

2.9 Patient handles

2.10 Leg supports

2.11 Function of the plastic handles

2.12 Colposcope attachment

2.13 IV stand

2.14 Proctology/rectoscopy attachment, assembly and operation

3. Maintenance & Cleaning

3.1 Routine check-ups

3.2 Cleaning

3.3 Lubrication schedule

3.4 Spare part list

3.5 Accessories list

4. Technical specification

4.1 Dimensions

4.2 Motors and controllers

4.3 Battery backup

4.4 Labels on the product

5. EU-Declaration of Conformity

6. Contact details

7. Functional control

FOREWORD

Welcome as the owner of a gynecology chair manufactured by Sjöbloms Sjukvårdsutrustning AB. We hope that you will be pleased and satisfied with our quality product.

If a problem should arise with any of our products, you are always welcome to contact us, and our service department will assist you with the expertise needed to solve the problem.

The user manual that you have in your hand contains all the information needed to use, maintain and service the product. The document is continuously updated as the products are modified. Therefore we reserve against possible changes in the user manual.

WARRANTY

Sjöbloms provides a two-year warranty for material, design and manufacturing defects, provided the products are not subjected to tampering or excessive wear. For the guarantee to apply, the product must have been used in accordance with the user manual, and the maintenance performed as specified.

IMPORTANT INFORMATION

This product is intended for professional medical use. The user manual should be read carefully before using the product, to avoid unnecessary risks for the patient and operator.

The user manual contains warnings where there is a risk of personal injury, which should be noted. Improper handling can be fatal.

A warning triangle is shown where particularly important information can be found.



Warning Triangle

1. FUNCTIONAL DESCRIPTION

This gynecology chair is meant for professional use in health care, for examination of the patient.

It is designed to be a functional and user-friendly tool for the operator and patient.



Note!

The product must only be used by competent health professionals who read the user manual, for correct handling.

General description

Gynaecology chair with electrical control of height and Trendelenburg position/seat section. Chromed or white chassis of welded steel pipes. The back rest is controlled with a gas spring or an electric motor. The cushions are stuffed with foamed plastic and upholstered with imitation leather.

The leg supports are available in two versions, they are attached to the accessory rail that runs along the seat cushion, which may also be used to attach certain accessories.

100mm center locked wire protected castors. Various accessories such as neck pillows, patient handles etc. are also available.

2. OPERATION


2.1 Connection of gynecology chair

The contact should be connected to a grounded 230V outlet. The gynecology chair is now ready for use.

2.2 Raising/lowering the height

Foot control

(the foot control is standard)



Follow the arrows to adjust the couch up and down.

Hand control

(the hand control is optional)



Ensure that no unauthorized persons are in the couch's working range, as a pinch hazard exists when raising and lowering the gynecology chair.


2.3 When the patient is being seated

For maximum safety, it is important that the staff help the patient to sit correctly in the seat. This also includes the placement of the actual gynecology chair.



Before examination:

1. Place the wheels in the direction shown in the picture.
2. Lock the central lock by pressing down the arc. Always keep the wheels locked during the examination!
3. Lower the seat to its lowest position. Help the patient to sit down in the middle of the seat cushion AND lean against the back support before you raise the chair.
4. When the patient is getting up again, it is important to grip the patient handles (optional, figure 4) and not the leg supports. The leg supports are not designed for this purpose.



Use caution when examining slightly heavier patients. For maximum balance of the chair, be sure to follow steps 1-4 above.

2.4 Adjusting the backrest

Hydraulic/gas spring

Grab at least one of the handles that are positioned on the side of the back rest.



Push the handle up and the gas spring helps to raise the back rest (variably adjustable).


Lowering the back rest

Push the handle in again and push the back rest down.



Ensure that no unauthorized persons are in the couch's working range, as a pinch hazard exists when raising and lowering the gynaecology chair.


Electrical, hand control (optional)



For gynecology chair with three motors.

Look at the symbols and follow the arrows to regulate the different functions of the gynecology chair.

Electrical, foot control (standard)




For gynecology chair with three motors.

The right pair regulates the back rest or the seat section/shock position.

Press the yellow button in the middle to switch between these functions.

The left pair adjusts the height of the gynecology chair.




2.5 Adjusting the seat section/Trendelenburg position

Electrical, hand control



Look at the symbols and follow the arrows to regulate the different

Electrical, foot control



For gynecology chair with three motors.

The right pair regulates the back rest or the seat section/shock position.


Press the yellow button in the middle to switch between these functions.

The left pair adjusts the height of the gynecology chair.



Ensure that no unauthorized persons are in the couch's working range, as a pinch hazard exists when raising and lowering the gynecology chair.


2.6 Replacing the paper roll



1. Turn the handle towards you.
2. Push the paper roll holder to the left, it will then detach on the right side, see figure 2.
3. Remove it entirely.
4. Slip on a new paper roll.
5. Replace it on the left side.
6. Replace the holder on the right side in the recessed groove.
7. Turn the handle again.

2.7 The function of the castors

The gynecology chair is equipped with central locking castors.



Unbraked position

2.8 Pull-out leg section


Pull the leg section towards you until it stops.



Lift so that the leg plate rises. Use caution; pinch hazard.



Press the underlying handle forward so that the leg plate's position locks against the seat pan. You can use your other hand for bracing



DONE!




NOTE!

The leg plate is only a support to rest your legs on. It is not designed for sitting. Damage caused by misuse are not covered by the warranty.



2.9 Patient handles



The patient handles can be adjusted along the accessory rail (arrow), and also be completely removed from the rail when necessary.

2.10 Leg supports

Standard




Integral




The gynecology chair is available with two different types of leg supports, Standard and Integral. Both models can be adjusted variously in all directions, up and down, forward and backward along the accessory rail.



2.11 Function of the plastic handles



The handles are tightened clockwise and eased counterclockwise. If you have to turn more than once, but you can not gain access because of the frame being in the way etc. as the picture two, pull the handle down (applies to handle A) or towards yourself (applies to handle B), then the tension mode is opened and you can turn back the handle without the achieved tension being released. When you then release the handle, it returns to the operational mode, and you can "repeat" your hold



2.12 Colposcope attachment



1. Remove the bolt and nut from the lift arm joint (A).
2. Remove the bolt and nut located on the colposcope (B).
3. Place the colposcope attachment bolt (B) through the lift arm joint and the extra bolt sent with the colposcopy attachment in the top hole (C). If the hole is not pre-drilled (on older chairs), drill a hole 65 mm above A.
4. Tighten the bolt and nut. The bolt in the top hole should be very tight. The bottom nut should not be too tight, so that the joints are not too hard to move.
5. Mount the colposcope onto the colposcopy attachment.
6. Insert the included top support in the top end of the colposcope (D).
7. Tighten the top support with the bolt (E).

2.13 IV stand




Remove the M10x55 bolt (A).

Mount the attachment as shown,
change bolt (A) to M10x60.

NOTE! Don't tighten too hard, since
the joint needs to move.

Mount the attachment around the
tube with the bolt (B).

2.14 Proctology/rectoscopy attachment, assembly instructions



Slide the supplied support legs of the procto/recto attachment under the gynecology chair from the short side.


Get assistance from someone and turn the gynecology chair so that it rests on the side.

Attach using the enclosed screws.



Proctology/rectoscopy attachment, user manual

Maximum recommended patient weight: 120 kilo.



- A. Lower the gynecology chair to its lowest position and remove the leg supports.
- B. Pull out the support legs.
- C. Adjust the feet so that **both** are resting solidly against the floor.
- D. For maximum stability the wheels should be directed as shown.



The attachment should look like this when it is fitted.



Guide in the square tube in the direction of the arrow.

The metal lock (see circle) automatically locks when you press in the procto/recto attachment. Only press the metal lock if necessary. Make sure that the attachment has been

Place the knee supports at the desired height. Ensure that both sides are fitted correctly before placing weight on the product.



The chrome plated patient handles can be adjusted along the accessory rail. Place them so that the pillow end up close to the seat edge.

Place pillows on each patient handle as shown in the figure.



3. MAINTENANCE AND CLEANING

3.1 Routine check-ups

Routine inspections should be made to ensure proper functionality.



Important!

Failure to perform the following routine can cause the patient and/or operator to be exposed to unnecessary risks and the couch may have a reduced life expectancy. Please check all functions and parts on a regular basis.

Check the following:

- Any unusual sounds.
- Gas spring functions.
- Castors and castor locking mechanisms.
- Any damage to electrical cables.
- Motor functions.
- That joints and fittings are lubricated according to the lubrication chart.

3.2 Cleaning

Cleaning artificial leather

Water and a mild soap solution is recommended.

Disinfection


If the product needs to be disinfected, use a diluted alcohol solution. Too strong an alcohol solution can have a drying effect and the imitation leather may deteriorate and crack.

Chlorine

Chlorine-base solutions are not recommended as a general rule, since it effects products and environment negatively, differently depending on concentration. If chlorine-based solutions are used, always wipe afterwards with damp cloth.

3.3 Lubrication scheme

Lubrication according to the schedule should be performed every 6 months or as needed.






1. First lock the castors.
2. Enlist the help of another person and tip the couch on the side, in order to lubricate the moving joints.

The spray can in the pictures above indicates where to lubricate. Use **"Smörjolja för Sjöbloms medicinska möbler"** or the equivalent for best results, article no. 5557001.

3.4 Spare part list gynecology chair


No.	Item no.	Description
1	G0001	Gas spring for back support
2	DET015-007CR	Paper roll holder
3	100321380T2LPZ	Castor standard
4	100601290T2LPV	White castors, central locking
5	343100+0015004A	Lift Motor Linak
6	J09980	Back support motor Linak
7	312100+1110004P	Trendelenburg motor Linak
8	FS2001	Foot control two motors, Linak
9	FS3001	Foot control three motors, Linak
10	HB824500004+11	Hand control two motors, Linak
11	HB834500003+11	Hand control three motors, Linak
12	WN55 S95-M12x25	Locking handles, large
13	WN55 S65-M8X16	Locking handles, small
14	DY-300-colour	Pads leg supports




3.5 Accessories



Neck cushion




Leg plate



Footrest



White castors



Colposcopes attachment




Lamp



Patient handles



Infusion stand



Proctology/rectoscopy attachment



Side supports, folding



Urology attachment



White-coated chassis

Other accessories/options

Battery backup, electrically adjustable back support.

3.6 Solutions for lubrication, cleaning and disinfecting

R-1, Cleaning solution for artificial leather

5557001, Lubrication

C068, Disinfection solution, DAX Plus



Cleaning solution

Lubrication Disinfection

4. TECHNICAL SPECIFICATION

4.1 Dimensions

Length: 130 cm

Length including leg plate: 173 cm

Width: 55 cm

Height: 53-98 cm

Length of backsupport: 90 cm

Angle of backsupport: +54°

Length of seat plate: 34 cm

Seat plate shock position angle 15 ° above the horizontal position, -10 ° below.

Weight: approx 55 kg

Lifting Capacity: 200 kg

4.2 Motors and controllers

The motors comply with EMC directives in accordance with standard EN 60601-1-2.

At full load, the motors can be used a maximum of 6 minutes/hour. The actuator manufacturer guarantees at least 10 000 cycles performance.

Controllers

Max. usage time 10% or 6 min/h at continuous operation. Complies with directives in accordance with standard EN 60601-1.

4.3 Battery backup

(optional)

General

The batteries used in the control unit are sealed lead batteries.

The included power cord should be plugged into an electrical outlet with 230 volts. The charger will recharge the battery in less than 12 hours. The charger can safely be connected for charging for extended periods.

If a battery is not used for a long time, it will gradually lose its stored energy. It is therefore recommended that the batteries are recharged at least 8 hours before usage.

Battery life

Battery life depends on several factors including the extent of discharging. If the battery is discharged too extensively between charges the lifespan will be reduced. Charging as often as possible is recommended.

The built-in buzzer "Charge the battery" that is built into the controller will sound when a discharge of 50% has occurred.


The life expectancy of the batteries is about 500-1200 charge cycles.

NOTE!



When the battery is depleted, the battery should be brought to a battery collection, to conserve the environment.


4.4 Labels on the products




The CE label contains the date, order number, serial number and our contact information.



The gynecology chair has a 200 kg lifting capacity.



Displays the adjustment of the back rest. There is a label on each side of the chair.



Displays the adjustment of the seat section/Trendelenburg position. There is a label on each side of the chair.

5. EU-Declaration of Conformity

EU DECLARATION OF CONFORMITY

according to MDD 2007/42/EC including Appendix
and according to CE-marking Directive 93/68/EEC

Type of equipment: Gynecology chair, electrically adjustable

Manufacturer: Sjöbloms Medical Equipment AB

Model: 015-2031 Product registration number: 6972

Model: 016-2031 Product registration number: 32719

The appendix contains documents that have been applied for verification of compliance.

Manufacturer

Sjöbloms

Myrängsvägen 8

S-891 50 ÖRNSKÖLDSVIK, SWEDEN

Sjöbloms Sjukvårdsutrustning AB ensures that the above products are in accordance with the EU Council MDD 93/42/EEC, as well as the Swedish legislation in force (1993:584) on medical devices and the regulations issued by Medical Products Agency (LVFS 2003:11).

Örnsköldsvik 2011-06-22



Karl-Gunnar Sjöblom

Production Manager

6. CONTACT INFORMATION

SJÖBLOMS

Myrängsvägen 8

891 50 Örnsköldsvik

Sweden

Tel: +46 660 105 80

Fax: +46 660 180 00

E-mail: info@sjobloms.com

Web page: www.sjobloms.com